

PRINCIPAL K. M. KUNDNANI COLLEGE OF PHARMACY

(GOVT. - AIDED, PCI & AICTE APPROVED, ACCREDITED by NBA and Affiliated to UNIVERSITY OF MUMBAI)

Tender No. 1

Tender Notice

Date: 24.4.2023

Sealed tenders are invited for HPLC with PDA detector with following minimum specifications

S. No.	Description	Qty
	HPLC system consists of:	1 set
1	Binary Gradient Pump Flexible expandability by using plug-in expansion bay system Flow rate range: 0.001 ~ 10.0 mL/min pH Range (recommended): 1.0-12.5 Operating Pressure: 4000 psi or more	1 no.
2	Photo Diode Array Detector Wavelength range: 200 - 800 nm (UV-Visible) Wavelength Accuracy: ±1nm Communication: USB2.0 or better Automatic WL Calibration using Hg lamp, D2 lamp or better Lamp Replacement: Front Access or better Lamp life time :2000 hrs (D2 lamp) or better Pressure limit: 1000 psi or better	1 no.
3	Manual Injector with standard loop	1 no.
4.	Chromatography System control software with 21 CFR Compliance & Integrating S/W	1 no.
5.	Required Tubing & Maintenance Kit	1 no.
6.	Bottle stand	1 no.
7.	Complete system provided with PC and printer optional	1 no.
8.	Optional	1 no.
	Auto sampler/ Quaternary Pump	1 no.
	Local HPLC Column oven	1 no.
	Column C18-5 μ (4.6mmID x 250mm), C8- 5 μ (4.6mmID x 250mm)	2 no.

Note: 1. Commercial quotation and technical specifications should be submitted separately.

2. Do not send quotation by fax or email.

3. The quotation should include the FOB, Air Freight, handling charges etc, (C.I.F. Mumbai) if applicable

The tender in sealed cover should be submitted on/ before 4.5.2023 addressed to the Principal, Prin. K. M. Kundnani College of Pharmacy, 23 Jote Joy Bldg, R. S. Marg, Cuffe Parade, Mumbai- 400 005. The Technical quotation for HPLC and Commercial quotation for HPLC should be marked on top of the sealed envelope.

U. Joshi
Principal,
Prin. K. M. Kundnani College of Pharmacy,

